



Federal Aviation Administration

Wheelchairs and mobility devices with lithium ion batteries, collapsible design, battery removed

Electric wheelchair, mobility scooter

"This description is for a wheelchair or mobility device that does not have a protective housing for its lithium ion battery. Lithium ion battery size is limited to 300 watt hours (Wh) or 25 grams equivalent lithium content (ELC). **(EW-01 Airline Approved Battery has 288 Wh)** One spare battery not exceeding 300 Wh or two spare batteries not exceeding 160 Wh (13.5 grams ELC) each may be carried in carry-on baggage.

EW-01 Travel Battery is a rechargeable Lithium battery

48 Volts and 6 amp

Lithium metal (non-rechargeable lithium) batteries are forbidden with these devices.

Lithium ion batteries must be removed from this type of mobility device and battery terminals protected from short circuit. **(EW-01 batteries are removable)** The lithium ion batteries must be carried in carry-on baggage only. The passenger must advise the airline of the battery location. The airline must notify the Pilot-in-command of the battery location in the cabin."